

Re: Tektronix 2230 Power Up Failure

Source: <http://sci.tech-archive.net/Archive/sci.electronics.repair/2008-01/msg01214.html>

- *From:* "Pierre-François" <pfmonet_NOSPAM@xxxxxxxxxxx>
 - *Date:* Sat, 26 Jan 2008 07:43:46 +0100
-

Oups! I'll power up mine these days to look at!

It has not been used since a while as I prefer to use the 2465b and keep the 2230 as spare...

If you find something in the battery area please let us know... I'll do the same...

pf

"dodger741" <dodger741@xxxxxxxxxxx> a écrit dans le message de news:
6bfaf4d6-30b7-4236-906c-e0692a6e91f2@xx

IIRC,2230 had a lithium coin cell holder on the top digital board,or a 123 lithium cell underneath the top digital board.(soldered to it)

Thanks for the quick response... Is the board you are referring to the one that has the plastic hinges for tilting the board up. Do you think the battery might be associated with the message I am getting as I suggested?
Roger

yes,it's the only board on the top of the instrument.
I don't know about the error messages.It's very possible.
Those batteries only have a 10 yr shelf life.
The 2230 has been out of support longer than that.

--I disassembled the unit and looked on both sides of the board. Nothing to be found that looks like a battery. There are two daughter boards attached to the back but they look like some sort of high frequency

Re: Tektronix 2230 Power Up Failure

analog boards because of the metal shield. Strange to think that there is a built in message that they don't mention in the manual! I may never know.

Jim Yanik

jyanik

at

kua.net